



ADAMS Accession no. ML22018A179

To: Lizette Roldan-Orteo

From: Dee Fairservis [dfairservis@energylab.com]
2393 Salt Creek Hwy.
Casper, WY 82601

Date: Wednesday, January 10, 2022

Subject: NRC Inspection Report 030-29502/2021-001: NOV

To whom it may concern,

The following are the corrective actions to NRC INSPECTION REPORT 030-29502/2021-001; AND NOTICE OF VIOLATION:

Due to the limited work/personal space within the radiochemistry lab, it became convenient to keep samples within the count room nearest to the instruments for loading purposes. This space includes analysts personal work desks where food and drink are consumed. Slowly, more and more samples that were to be counted made their way into the count room and it started moving into the area of the work desks.

In order to correct this potential health hazard, samples will no longer be allowed to be stored in the count room. The only samples in the count room will be samples that are in each instrument counting. Samples were removed from the count room on 12/30/21.

In addition, a tape line was placed on the floor to separate the areas where food will be stored and where instruments are located. Samples will be prohibited to cross over this line into the clerical area. The tape was placed on the floor on 08/17/2021 the day after the NRC audit.

Additional follow up with laboratory analysts will be in place to ensure that these changes remain in place. In closing we hope that the corrective action for this NOV will rectify these findings.

Regards,

Dee Fairservis, RSO