

**From:** [Gallagher, Robert](#)  
**To:** [Chris Beaulieu](#)  
**Cc:** [Patel, Netra](#)  
**Subject:** Diagnostic Imaging of Southbury, LLC Request for Additional Information, Control No. 629848  
**Date:** Monday, March 7, 2022 9:47:00 AM  
**Attachments:** [image003.png](#)

---

Licensee: Diagnostic Imaging of Southbury, LLC  
License No.: 06-31024-01  
Docket No.: 030-36882  
Control No.: 629848

PLEASE REPLY TO THIS EMAIL TO CONFIRM RECEIPT OF THIS REQUEST FOR ADDITIONAL INFORMATION

Dear Mr. Beaulieu:

This is in reference to your application dated January 10, 2022, requesting termination of NRC License No. 06-31024-01. In order to continue our review, please provide the following additional information:

1. Please provide the last date when the radioactive material was in use.
2. The application states that the radioactive materials were transferred to Eckert and Ziegler. Please provide confirmation documentation of receipt for the radioactive materials.
3. Your application did not specify the instrument used to measure the area contamination wipes. Please specify your instrumentation that was used and documentation of calibration.
4. Please provide a facility diagram that includes all the location where the radiological surveys were performed.
5. The documentation of radiological surveys in your termination request appear to show that all of the instruments used were out of calibration at the time the surveys were performed. Please provide documentation of radiological surveys performed with properly calibrated instrumentation.

In order to continue prompt review of your application, we request that you submit your response to this email within 30 calendar days from the date of this email and reference mail control number 629848.

An electronic version of the NRC's regulations is available on the NRC Web Site at: [www.nrc.gov](http://www.nrc.gov). Additional information regarding medical uses of radioactive materials may be obtained on the NRC Web Site at: <http://www.nrc.gov/materials/miau/med-use-toolkit.html>. This site also provides the updated Training and Experience NRC Form 313A series of forms and guidance, as well as information on the revised regulations for naturally-occurring and accelerator-produced radioactive materials (NARM).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web Site at: <http://www.nrc.gov/reading-rm/adams.html>. Please be aware that you may request that certain portions of your submittal to NRC be withheld from public disclosure as proprietary information. To do this, you must execute an affidavit as specified in 10 CFR 2.390. You must list all portions that you wish to be held proprietary, along with your reasoning as to why that is appropriate. While it is allowable, please refrain from submitting proprietary information in support of a license unless necessary. Keep in mind that all NRC licenses are considered to be in the public domain, and therefore may be viewed by any member of the public who requests to see them.

If you have any questions regarding this request for additional information, please contact me via electronic mail at [robert.gallagher@nrc.gov](mailto:robert.gallagher@nrc.gov).

Thank you for your cooperation.

Regards,

Robert L. Gallagher  
Health Physicist  
Medical & Licensing Assistance Branch  
U.S. NRC, Region I  
2100 Renaissance Blvd.  
King of Prussia, PA 19406  
(610) 337-5182 office  
(610) 337-5269 fax

