



TRI STATE MATERIALS TESTING LAB LLC.

New England Regional Office

60 Woodlawn Road, Berlin CT 06037 • Tel: 203-949-7733 • Fax: 203-949-7735

April 20, 2022

Jason vonEhr
Mail Control No. 630237 USNRC
Region 1
Division of Radiological Safety and Security 2100
Renaissance Boulevard
King of Prussia, PA 19406

SUBJECT: TRI STATE MATERIALS TESTING LAB, LLC, REQUESTED ADDITIONAL INFORMATION,
MAIL CONTROL NO. 630237

Dear Mr. VonEhr:

This is in reference to our application dated February 8, 2022, requesting to renew NRC License No. 06-31462-01. We are providing the requested additional information.

1. No response is required.
2. Our present radioactive material authorizations are sufficient.
3. No response is required.
4. Before using licensed materials, authorized users will have successfully completed one of the training courses described under "Criteria" in the section titled "Training for Individuals Working in or Frequenting Restricted Areas" in NUREG-1556, Vol 1, Revision 2.
5. We will either possess and use, or have access to and use, a radiation survey meter that meets the criteria in the section titled "Radiation Safety Program—Radiation Monitoring Instruments" in NUREG—1556, Volume 1, Revision 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses," in the event of an incident.
6. We will implement and maintain the operating, emergency, and security procedures in Appendix G to NUREG—1556, Volume 1, Revision 2, "Consolidated Guidance About Materials



TRI STATE MATERIALS TESTING LAB LLC.

New England Regional Office


60 Woodlawn Road, Berlin CT 06037 • Tel: 203-949-7733 • Fax: 203-949-7735

Licenses: Program-Specific Guidance About Portable Gauge Licenses.” Copies of these procedures will be provided to all gauge users and will be available at each jobsite.

7. Leak tests will be performed at intervals approved by the NRC or an Agreement State and specified in the Sealed Source and Device registration certificate. Leak tests will be performed by an organization licensed by the NRC or an Agreement State to provide leak testing services to other licensees; or by using a leak test sample collection kit supplied by an organization licensed by the NRC or an Agreement State to provide leak test kits and/or sample analysis services to other licensees and according to the kit supplier's instructions. Records of leak test results will be maintained.
8. a) Routine cleaning and lubrication: We will implement and maintain procedures for routine maintenance of our gauges according to each manufacturer's written recommendations and instructions.
b) Nonroutine maintenance or repair operations that require detaching the source or source rod from the gauge: The gauge manufacturer, or other person licensed by the NRC, or an Agreement State will perform nonroutine maintenance or repair operations that require detaching the source or source rod from the gauge.

Sincerely,

Tri State Materials Testing Lab LLC


Habib A. Chaudhary
Principal

TSMT

www.tristate-testing.com